

OOH

·S A L U D·

EQUIPO DE PROTECCIÓN PERSONAL

KN95 Respirator Face Mask



Cinco capas de filtración reforzada



Alta capacidad de filtración



Diseño 3D para mayor comodidad y mayor ajuste

Información de producto

Certificación: FDA + CE

Código de producto FDA: LYU, MSH

Estándar de producto: GB2626-2006

Rendimiento del filtro: 95%+

Medidas de embalaje:

Unidades por caja: 30

Cajas por cartón master: 40

Unidades totales por cartón master: 1,200

Dimensiones por cartón master: 59x54x55cm

Peso bruto por cartón master: 9,5kg

Peso neto por cartón master: 8,8k

Cartones master por pallet: 18

Pallet: 21,600 unidades

Tamaño y dimensiones del Pallet:

Talla L: 120cm x W 100 cm x H 141 cm

Peso total: 180 KG

Órdenes

Orden de compra mínima:

21,600 unidades (1 pallet)

Ventas en múltiplos de 21,600 unidades

Envíos globales

Inventario disponible:

300,000 unidades a embarcar de inmediato

400,000 unidades adicionales a embarcar 17/4/2020

500,000 unidades adicionales a embarcar 24/4/2020

4-7 días de tránsito aéreo

*Opción de asegurar carga disponible

Condiciones de pago:

50% pago por adelantado


50% pago pre-embarque



Imagen de la caja de KN95

- * Cada envío es inspeccionado por nuestro equipo de control de calidad en nuestras oficinas en Shanghai, China
- * Prueba de vida disponible bajo solicitud

PROBABILIDAD DE CONTAGIO



MUY ALTA



ALTA



MEDIA



BAJA

TECHNICAL DATA SHEET KN95 Respirator Face Mask				
Use: for Paricles such as those from grinding, sanding, sweeping, sawing, bagging, or processing minerals, coal, iron ore, flour, metal, wood, pollen, and certain other substances. Liquid or non-oil based particles from sprays that do not also emit oil aerosols or vapors.				
Do not use for: Gas and Vapors, oil aerosols, asbestos, arsenic cadmium, lead, 4,4' - methylene dianiline (MDA) or sandblasting. This respirator dues not supply oxygen.				
Time Use Limitation: If respirator becomes damaged, soiled, or breathing becomes difficult, leave the contaminated area immediately and replace the Respirator. The suggested duration is 3-4 hours.				
Characteristic	Specification			
Material Composition	Mask body	Outer Layer	Non-woven fabric	
		Filter Layer	Meltblown fabric	
		Inner Layer	Non-woven fabric	
	Nose clip	Aluminum		
	Elastic band	Spandex / Nylon		
Description	KN95 Respirator Mask (Non-Medical)			
Color	Mask	White		
	Elastic band	White		
Nose Clip Style	Flat Nose	Clip		
Dimension	Body Size	Length	135	mm
		Width	130	mm
		Height	60	mm
	Nose Clip	Length	90	mm
	Elastic band	Length	100	mm
Storage Condition			Country of Origin	
Store under cool, clean, and dry conditions. Avoid excessive heat (over 40°C or 104°F)			China	
List of items				
Item Code	Description		Packaging Information	
HM-KN95	KN95 Respirator Mask		30 Respirators/Box, 40 Boxes/Case, 1200 Respirators/Case	







中国认可
国际互认
TESTING
CNAS L6780



TEST REPORT



扫一扫关注我们

扫一扫查询真伪

YJ202004067

报告编号
REPORT NO.

国纺委字第 YJ202004067 号

产品名称
NAME OF SAMPLE

KN95 口罩

委托单位
CUSTOMER

检验类别
TEST CATEGORY

委托检验

浙江省轻工业品质量检验研究院

(浙江省纺织测试研究院)

Zhejiang Light Industrial Products Inspection and Research Institute

国家纺织服装产品质量监督检验中心(浙江)

National Textiles and Garment Quality Supervision Inspection Center(Zhejiang)

浙江省轻工业品质量检验研究院
国家纺织服装产品质量监督检验中心(浙江)

检验报报

国纺委字第 YJ202004067 号

第 1 页 共 3 页

委托单位名称 Name of Customer		地址 Address	---
生产单位 Manufacturer		地址 Address	---
样品信息 Sample information	样品名称 Name of sample: KN95 口罩 样品特性 Characteristics: 白色 商标 Trademark: --- 规格/号型 Specification/model: HM-KN95 等级 Level: KN95 安全技术类别 Category of safety specification: --- 样品款号/货号 Art. No.: ---		
以上为客供信息 (Above-mentioned information by Customer-supplied)			
来样方式 The sent way of sample	快递	样品数量 Sample quantity	20 只
送检日期 Receiving Date of Sample	2020-03-19	检测类别 Test Category	委托检验
判定依据 Rating Requirements	GB 2626-2006		
检测结论/Test Summary: 实测结果详见附页。			
 (检验报报专用章) Test Seal 批准日期/ Date of Approval: 2020-03-19			
备注 Remarks	样品未经预处理。		

签发:
Approved by

张雪芳

INFORME DE CALIDAD

检验报报

国纺委字第 YJ202004067 号

第 2 页 共 3 页

序号	检测项目	检测方法	单位	标准要求 (KN95)	实测值	单项评价	结果备注
1	过滤效率	GB 2626-2006 6.3	%	≥95.0	99.0	符合	——
2	呼吸阻力	GB 2626-2006 6.5 6.6	Pa	≤350	102	符合	——
	呼气阻力		Pa	≤250	78		

检验报报

国纺委字第 YJ202004067 号

第 3 页 共 3 页

样品照片



—以下空白—

INFORME DE CALIDAD

声明

- 一、本机构保证及测的公正性、独立性和诚实性，对及测结果负责，对委托方所提供的及测样品保密和保护所有权。
- 二、本报告无签发人员签字或未加盖本机构红色及验及测专用章无效。
- 三、报告涂改无效。复印件未重新加盖本机构红色及验及测专用章无效。
- 四、本报告的及测数据和结果仅对送及样品负责。
- 五、本报告各页均为报告不可分割之部分，使用者单独抽出某些页导致误解或用于其它用途由此造成的后果，本机构不负相应的法律责任。
- 六、委托方若对本报告有异议，应研时向本机构提出。政府行政管理部门下达的指令性任务，被及方对及验结果有异议时，应按政府行政管理部门文件规定研国家相关法律、法规规定进行。

DECLARATION

1. Our organization guarantees impartiality, independence and honesty of inspection, and is responsible for the results of inspection, keeping the samples supplied by the entrusting party confidential and at the same time protecting the ownership of the samples supplied.
2. The test report will be deemed invalid without signatures of the inspector/reviewer and authorized personnel, and the red special inspection stamp of our organization.
3. The test report will be invalid if it is altered. Copies of the report are invalid without the red special inspection stamp of our organization.
4. The test results shown in this report is only valid for the tested sample.
5. All the pages of the report are integral parts of the report. Our organization will not be responsible for any misunderstanding or other results caused by using separate page(s) of the report.
6. If there is any dissent of the report, the entrusting party shall notify our organization timely. For the mandatory inspection given by governmental administration departments, and dissent about the sample being tested or test results on the report should be dealt with in accordance with national regulations.

浙江省轻工业品质量及检验研究院研究设的及验中设 The Affiliated Inspection Centers

浙江省轻工业品质量及检验研究院 Zhejiang Light Industrial Products Inspection and Research Institute 国家纺织服装产品质量监督及验中设（浙江） National Textiles and Garment Quality Supervision Inspection Center (Zhejiang) 地址：浙江省杭州市江干区下沙路300号6号楼 Address: Building No.6, 300 XiaSha Road, Hangzhou, Zhejiang	联系电话：0571-85122669 Telephone: 0571-85122669 E-Mail: 1137907994@qq.com
国家家具产品质量监督及验中设（浙江） National Center for Quality Supervision Inspection of Furniture (Zhejiang) 浙江省室内安全家具产品质量及验中设 Zhejiang Center of Quality Test for Indoor Safety and Furniture Products 地址：浙江省杭州市余杭区良渚街道经一路1号良渚大学科技园4号楼 Address: Building 4 LiangZhu University Science and Technology Park, No.1, JingYi Rd., Yuhang District, Hangzhou, Zhejiang	联系电话：0571-89009556 Telephone: 0571-89009556 E-Mail: 2047699564@qq.com
浙江省轻工研五金产品质量及验中设 Zhejiang Center of Quality Test for Light Industry and Hardware Products 浙江省体育用品质量及验中设 Zhejiang Center of Quality Test for Sports Products 地址：浙江省杭州市西湖区天目山路222号3号楼 浙江省杭州市余杭区良渚街道经一路1号良渚大学科技园3号楼 Address: No. 222 Tianmushan Rd., Hangzhou, Zhejiang Building 3 LiangZhu University Science and Technology Park, No.1, JingYi Rd., Hangzhou, Zhejiang	联系电话：0571-89001107 Telephone: 0571-89001107 E-Mail: wujtest@126.com
国家锁具产品质量监督及验中设（浙江） National Center for Quality Supervision Inspection of Lock (Zhejiang) 浙江省锁具产品质量及验中设 Zhejiang Center of Quality Test for Lock Products 地址：浙江省杭州市塘南南路24号 Address: No. 24 Tangmiao Rd., Hangzhou, Zhejiang	联系电话：0571-85027738 Telephone: 0571-85027738 E-Mail: Locktest@126.com

网上业务受理/报告查询: <http://www.zjttj.cn:807>
Online Business-Reception/Report Inquires: <http://www.zjttj.cn:807>

投诉电话: 0571-85023552
Complaint Tel: 0571-85023552

APROBACIÓN FDA

FDA

Medical Device Registration

The 3rd Party Certificate of



Note:

This file is Not being issued by FDA. WWW.GTS as the 3rd party, produce it, intended to facilitate customer display & transmit information. The following contents, FDA registered Facility/Owner/Operator&FDA listing Medical Device, are excerpted from database at www.fda.gov.

Establishment:

322204

Registration Number / FEI Number*:

* Firm Establishment Identifier (FEI) should be used for identification of entities within the imports message set

Status: **Active**

Date of Registration Status: 2020

Owner/Operator

322204

Owner/Operator Number: [10064332](#)

Devices Listing Information

Proprietary Name	Product Codes	Device Class	Listing Number	Establishment Operations
mask	LYU	1	D37****	Manufacturer
KN95 Mask	LYU	1	D37****	Manufacturer
KN95 Mask	MSH	1	D37****	Manufacturer
N95 Mask	MSH	1	D37****	Manufacturer

Please careful protect your Listing Number.

Professional FDA Registration Services, byShanghai Global Testing Services Co., Ltd.

More details on the website: <http://www.gts-lab.com>

Need help? Contact us, GTS, at +86(021) 33637866 & info@gts-lab.com